

FUGRO TECHNICAL SERVICES LIMITED

Fugro Development Centre,
5 Lok Yi Street, Tai Lam,
Tuen Mun, N.T.,
Hong Kong.

Tel : +852 2450 8233
Fax : +852 2450 6138
E-mail : matlab@fugro.com
Website : www.fugro.com

MaterialLab

Report No. : 181004EN180628(7)



Page 1 of 1

Test Report of Chemical Analysis of Paint

Information Supplied by Client

Client : Hong Kong Niceview International Ltd.
Client's address : Flat H, 21/F, Golden Dragon Ind. CTR., Block III, 172-180 Tai
Lin pak Road, Kwai Chung, N.T.
Project : HKGCB Green Building Product Labelling Scheme
Sample description : Niceview Exterior Paint
Sample identification : Niceview Exterior Paint
Test required : Determination of phthalates in the product by GC/MSD
- Benzyl butyl phthalate (BBP)
- Dibutyl phthalate (DBP)
- Di-(2-ethylhexyl) phthalate (DEHP)
- Diisononyl phthalate (DINP)
- Di-n-octyl phthalate (DNOP)
- Diisodecyl phthalate (DIDP)

Laboratory Information

Lab. sample ID. : EN180628/1
Date of receipt of sample : 10/03/2018
Date test commenced : 16/04/2018
Date test completed : 19/04/2018
Test method used : In-housed method based on method CPSC-CH-C1001-09.3

Result:

Analytes	Concentration, % (w/w)
Benzyl butyl phthalate (BBP)	< 0.1
Dibutyl phthalate (DBP)	< 0.1
Bis-(2-ethylhexyl) phthalate (DEHP)	< 0.1
Di-n-octyl phthalate (DINP)	< 0.1
Diisononyl phthalate (DNOP)	< 0.1
Diisodecyl phthalate (DIDP)	< 0.1

Tested by : M.L.Yung, H.Y. Li

Certified by : 
Approved Signatory: HO Kin Man, John
Manager – Chemistry Department

Date : 8/5/2018

** End of Report **

Note : This report refers only to the sample(s) tested.

1